

This medical device is intended for patient positioning. This is not a restraint system.  
Before each use, it is important to check: the presence of the label, the stitches of the straps: that they do not fray or break, that the device does not show tears, snags, holes or signs of fraying (fabric and straps), that the buckles are not cracked or have no weak fastening.


## WALKING AID



Equipped with a self-centering  
50 mm buckle

Composition :  
combination of polyurethan foam  
and jersey/cotton polyester fabric  
Jersey/PU coating  
Polyester strap  
100% polyester fabric

This belt secures the patient. It is used as a walking aid.



CADM2-PF



The use of the device requires the presence and vigilance of a third person. When used by an agitated or disoriented person, monitoring becomes imperative.  
This medical device must be implemented or trained by a health professional.



v 05.02/23